

OCT 21 2005

K052609

**CD HORIZON® Spinal System  
Summary of Safety and Effectiveness  
September 2005**

**I. Company: Medtronic Sofamor Danek, Inc. USA  
1800 Pyramid Place  
Memphis, TN 38132  
(901) 396-3133**

**Contact: Richard W. Treharne, PhD  
Senior Vice President Regulatory Affairs**

**II. Proposed Proprietary Trade Name: CD HORIZON® Spinal System**

**III. Classification Name(s)/Product Code(s): Spinal Interlaminar Fixation and Spinal Intervertebral Fixation Orthosis and/or Pedicle Screw Spinal System (per 21 CFR Section 888.3050, 888.3060 and/or 888.3070) Product Codes: MNI, MNH, KWP, KWQ, NQP and NKB**

**IV. Product Description**

The CD HORIZON® Spinal System consists of a variety of rods, hooks, screws, CROSSLINK® plates, staples, and other connecting components used to build a spinal construct. Instrumentation is also available to facilitate implantation of the device components.

The CD HORIZON® Spinal System is intended to help provide immobilization and stabilization of spinal segments as an adjunct to fusion of the thoracic, lumbar, and/or sacral spine. The CD HORIZON® Spinal System implant components can be rigidly locked into a variety of configurations, with each construct being tailor-made for the individual case. If necessary, the CD HORIZON® Spinal System can be connected to the VERTEX™ Reconstruction System through a rod connector.

Certain implant components from other Medtronic Sofamor Danek spinal systems can be used with the CD HORIZON® Spinal System. These components include TSRH® rods, hooks, screws, plates, CROSSLINK® plates, connectors, staples and washers; GDLH® rods, hooks, connectors and CROSSLINK® bar and connectors; LIBERTY® rods and screws; DYNALOK PLUS® bolts; and Medtronic Sofamor Danek Multi-Axial rods and screws.

CD HORIZON® hooks are intended for posterior use only. CD HORIZON® staples and CD HORIZON® ECLIPSE® rods and screws are intended for anterior use only. However, for patients of smaller stature, CD HORIZON® 4.5mm rods and associated components may be used posteriorly.

The purpose of this 510(k) submission is to add sterile PEEK rods to the CD HORIZON® Spinal System.

page 1 of 3

000105

## V. Indications

The CD HORIZON® Spinal System is intended for posterior, non-cervical fixation for the following indications: degenerative disc disease (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies); spondylolisthesis; trauma (i.e., fracture or dislocation); spinal stenosis; curvatures (i.e., scoliosis, kyphosis and/or lordosis); tumor; pseudarthrosis; and/or failed previous fusion.

When used in a percutaneous, non-cervical, posterior approach with the SEXTANT instrumentation, the CD HORIZON® screws are intended for the following indications: degenerative disc disease (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies); spondylolisthesis; trauma (i.e., fracture or dislocation); spinal stenosis; curvatures (i.e., scoliosis, kyphosis and/or lordosis); tumor; pseudoarthrosis; and/or failed previous fusion.

Except for hooks, when used as an anterolateral thoracic/lumbar system, CD HORIZON® components such as ECLIPSE® components are intended for the following indications: (1) degenerative disc disease (as defined by back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies), (2) spinal stenosis, (3) spondylolisthesis, (4) spinal deformities (i.e., scoliosis, kyphosis, and/or lordosis), (5) fracture, (6) pseudarthrosis, (7) tumor resection, and/or (8) failed previous fusion.

The CD HORIZON® SPIRE Plate is posterior, non-pedicle supplemental fixation device, intended for use in the non-cervical spine (T1 – S1). It is intended for plate fixation/attachment to spinous process for the purpose of achieving supplemental fusion in the following conditions: degenerative disc disease - defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies; spondylolisthesis; trauma (i.e., fracture or dislocation); and/or tumor.

The CD HORIZON® LEGACY 3.5mm rod and associated components, when used as a pedicle screw fixation system of the non-cervical posterior spine in skeletally mature patients, are indicated for one or more of the following: (1) degenerative spondylolisthesis with objective evidence of neurologic impairment, (2) fracture, (3) dislocation, (4) scoliosis, (5) kyphosis, (6) spinal tumor, and/or (7) failed previous fusion (pseudarthrosis).

In addition, when used as a pedicle screw fixation system, the CD HORIZON® LEGACY 3.5mm rod and associated components, are indicated for skeletally mature patients: (a) having severe spondylolisthesis (Grades 3 and 4) of the fifth lumbar-first sacral (L5-S1) vertebral joint; (b) who are receiving fusions using autogenous bone graft only; (c) who are having the device fixed or attached to the lumbar and sacral spine (L3 and below); and (d) who are having the device removed after the development of a solid fusion mass.

When used as a pedicle screw system in skeletally mature patients, the CD HORIZON® Spinal System PEEK rods and associated components are intended to provide immobilization and stabilization of spinal segments as an adjunct to fusion in the treatment of the following acute and chronic instabilities of the thoracic, lumbar and sacral spine: (1) degenerative spondylolisthesis with objective evidence of neurologic impairment, (2) kyphosis, and/or (3) failed previous fusion. Additionally, when used as a pedicle screw device, the CD HORIZON® Spinal System PEEK rod constructs are indicated for use in patients who: (1) are receiving fusion with autogenous graft only, (2) who are having the device attached to the lumbar or sacral spine, and/or (3) who are having the device removed after the development of a solid fusion mass.

In order to achieve additional levels of fixation, the CD HORIZON® Spinal System rods may be connected to the VERTEX™ Reconstruction System with the VERTEX™ rod connector. Refer to

K052609

the VERTEX™ Reconstruction System Package Insert for a list of the VERTEX™ indications of use.

**VI. Substantial Equivalence**

Documentation, including a sterilization validation report, was provided which demonstrated the subject rods to be substantially equivalent to predicate CD HORIZON® PEEK components previously cleared in K050809 (SE 06/14/05).

page 3 of 3

000107



Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

OCT 21 2005

Richard W. Treharne, Ph.D.  
Senior Vice President, Regulatory Affairs  
Medtronic Sofamor Danek, Inc.  
1800 Pyramid Place  
Memphis, Tennessee 38132

Re: K052609  
Trade/Device Name: CD Horizon Spinal System  
Regulation Number: 21 CFR 888.3070  
Regulation Name: Pedicle screw spinal system  
Regulatory Class: III  
Product Code: NKB, NQP, KWP, KWQ, MNH, MNI  
Dated: September 21, 2005  
Received: September 22, 2005

Dear Dr. Treharne:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Page 2 – Richard W. Treharne, Ph.D.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0120. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,

A handwritten signature in black ink, appearing to read "Mark N. Melkerson". The signature is written in a cursive style and is positioned above the typed name.

Mark N. Melkerson  
Acting Director  
Division of General, Restorative,  
and Neurological Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

510(k) Number (if known): K052609

Device Name: CD HORIZON® Spinal System

**Indications for Use:**

The CD HORIZON® Spinal System is intended for posterior, non-cervical fixation for the following indications: degenerative disc disease (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies); spondylolisthesis; trauma (i.e., fracture or dislocation); spinal stenosis; curvatures (i.e., scoliosis, kyphosis and/or lordosis); tumor; pseudarthrosis; and/or failed previous fusion.

When used in a percutaneous, non-cervical, posterior approach with the SEXTANT instrumentation, the CD HORIZON® screws are intended for the following indications: degenerative disc disease (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies); spondylolisthesis; trauma (i.e., fracture or dislocation); spinal stenosis; curvatures (i.e., scoliosis, kyphosis and/or lordosis); tumor; pseudoarthrosis; and/or failed previous fusion.

Except for hooks, when used as an anterolateral thoracic/lumbar system, CD HORIZON® components such as ECLIPSE® components are intended for the following indications: (1) degenerative disc disease (as defined by back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies), (2) spinal stenosis, (3) spondylolisthesis, (4) spinal deformities (i.e., scoliosis, kyphosis, and/or lordosis), (5) fracture, (6) pseudarthrosis, (7) tumor resection, and/or (8) failed previous fusion.

The CD HORIZON® SPIRE Plate is posterior, non-pedicle supplemental fixation device, intended for use in the non-cervical spine (T1 – S1). It is intended for plate fixation/attachment to spinous process for the purpose of achieving supplemental fusion in the following conditions: degenerative disc disease - defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies; spondylolisthesis; trauma (i.e., fracture or dislocation); and/or tumor.

**510(k) Number:** The CD HORIZON® LEGACY 3.5mm rod and associated components, when used as a pedicle screw fixation system of the non-cervical posterior spine in skeletally mature patients, are indicated for one or more of the following: (1) degenerative spondylolisthesis with objective evidence of neurologic impairment, (2) fracture, (3) dislocation, (4) scoliosis, (5) kyphosis, (6) spinal tumor, and/or (7) failed previous fusion (pseudarthrosis).

In addition, when used as a pedicle screw fixation system, the CD HORIZON® LEGACY 3.5mm rod and associated components, are indicated for skeletally mature patients: (a) having severe spondylolisthesis (Grades 3 and 4) of the fifth lumbar-first sacral (L5-S1) vertebral joint; (b) who are receiving fusions using autogenous bone graft only; (c) who are having the device fixed or attached to the lumbar and sacral spine (L3 and below); and (d) who are having the device removed after the development of a solid fusion mass.

When used as a pedicle screw system in skeletally mature patients, the CD HORIZON® Spinal System PEEK rods and associated components are intended to provided immobilization and stabilization of spinal segments as an adjunct to fusion in the treatment of the following acute and chronic instabilities of the thoracic, lumbar and sacral spine: (1) degenerative

*Spinal, Back, Fusion Sign-Off*  
vision of General, Restorative,  
and Neurological Devices

*K052609*

0001

K052609

spondylolisthesis with objective evidence of neurologic impairment, (2) kyphosis, and/or (3) failed previous fusion. Additionally, when used as a pedicle screw device, the CD HORIZON® Spinal System PEEK rod constructs are indicated for use in patients who: (1) are receiving fusion with autogenous graft only, (2) who are having the device attached to the lumbar or sacral spine, and/or (3) who are having the device removed after the development of a solid fusion mass.

In order to achieve additional levels of fixation, the CD HORIZON® Spinal System rods may be connected to the VERTEX™ Reconstruction System with the VERTEX™ rod connector. Refer to the VERTEX™ Reconstruction System Package Insert for a list of the VERTEX™ indications of use.

Prescription Use   X    
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use         
(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

---

Concurrence of CDRH, Office of Device Evaluation (ODE)

*Janet Melvin for MXM*  
**(Division Sign-Off)**

**Division of General, Restorative,  
and Neurological Devices**

**510(k) Number** K052609

*page 2 of 2*

000180